

3M Medical Solutions Division

3M Center

St. Paul, MN 55144-1000

Certificate of Analysis

Product Identification

3M Catalog Number	9680
Product Description	Single Use Blade Assembly
3M Stock Number	70-2011-8507-4
3M SAP Number	7100213055
3M Lot Number	20241125
Date of Manufacture	11/25/24

Test Data:

Test	Method	Specification	Results
Cut Performance	Per EQB-044	40 pcs of tynex in 3 locations on blade	Pass
Blade Assembly	Per SOP042	Must withstand 15 lb shear test	Pass
Comb Flatness	Per Drawing	< 0.025 mm	Pass
Blade Flatness	Per Drawing	< 0.03 mm	Pass

This Letter Certifies that the product listed on this certificate has:

- A) Been manufactured in accordance with Good Manufacturing Practices for Medical Device Manufacturers as specified by the Department of Health and Human Services FDA QSR Federal Register 21 CFR Part 820.B, and ISO 13485; 2003.
- B) Been manufactured in accordance with 3M Company Specifications

Prepared By: Tom Baumgard

QA/QC Representative: Tom Baumgard

Date: 12/9/24