

Descripción del Producto	Elemento de protección personal que cubre boca y nariz, elaborado con Polipropileno no tejido, resistente a fluidos, antiestático, con ajustes elásticos a la cabeza. Desechable.
Función	Eficaz barrera física que protege contra salpicaduras, suciedad y polvo. La protección se realiza tanto para el usuario como para el ambiente ya que, en el caso de enfermedad del usuario, los microorganismos quedan en el filtro de la mascarilla y no en el ambiente.
Material	Tejido exterior: Polipropileno no tejido Capa de filtro: Tela de polipropileno para pulverizar Forro: Polipropileno no tejido
Almacenamiento	Lugar fresco y seco a temperatura entre 5° y 30 ° C. Evite luz solar directa
Vida útil:	3 años



DIMENSIONES	
Largo	17.3 ± 0.5 Cm
Ancho	9.5 ± 0.5 Cm
Peso	Por cada capa 20g/cm² mínimo

Empaque	
Bolsa	50 piezas
Caja	60 Bolsas
Caja	50 x 40 x 55 Cm
Caja	Peso Bruto 12.5 Kg

Certificate of Compliance



No. 0H200327M.SRM0380

Certificate's
Holder:

Suzhou Ruihe Medical Science &
Technology Company
Xinle Village of Foreign Merchant Development
Zone, Zhenze Town, Wujiang District, Suzhou City,
Jiangsu Province, China

Certification ECM
Mark:



Product:
Model(s):
Verification to:

Disposable Protective Mask
RH-P001
Standard:
EN 149:2001+A1:2009
related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary certification of products. RG01_ECM rev3, available at: www.entecertma.it

Issuance date: 27 March 2020

Expiry date: 26 March 2025

Reviewer
Technical expert
Amanda Payne



Approver
ECM Service Director
Luca Bedonni

