





检验检测报告 TEST REPORT



STFWT20207229

Product Name

KN95 Mask

Trust Unit

K-Captain (Jinhua) Protection Technology Co., Ltd.

Manufacturer

K-Captain (Jinhua) Protection Technology Co., Ltd.

Test Category

Trust Test



Test Report

| 07229 | 2 | Page 1 of 5 | |
|---|---|---|--|
| KN95 Mask | Specification Type | (18) | |
| | Trademark | K-Captain | |
| K-Captain (Jinhua) Protection Technology Co., Ltd. | Tel | 13405784110 | |
| K-Captain (Jinhua) Protection Technology Co., Ltd. | Sample Grade | KN95 | |
| 40 | Sample Receiving Date | 2020-03-31 | |
| Trust Test | Serial Number | | |
| Accord with detecting request | 600 | 63 | |
| GB 2626-2006 《Respiratory protective | e equipment. Non-powere | d air-purifying particle respirator | |
| | K-Captain (Jinhua) Protection Technology Co., Ltd. K-Captain (Jinhua) Protection Technology Co., Ltd. 40 Trust Test Accord with detecting request | KN95 Mask Trademark K-Captain (Jinhua) Protection Technology Co., Ltd. K-Captain (Jinhua) Protection Technology Co., Ltd. Sample Grade Sample Receiving Date Trust Test Serial Number Accord with detecting request | |

SignatuimDate: 2020-04-08

Remarks

Conclusion

Test items marked"——"are not Subject to or not meet the conditions for individual judgment.

the items inspected and does not mean that the uninspected items or functions meet the requirements.

The results apply to the sample as received.

Approver

群.独

standard KN95 level.

Examiner

京京市

Majortester

茶蕊文

| ~ | | 1 | | | Pag | ge 2 | of 5 | |
|--------|------------------------|----------|---|-------------------------|------------|-------------------------------|-----------|-------|
| Serial | Test Items | Res | | Individ: Judgme | | | | |
| 1 | General | 88 | a) Materials: a) Materials in direct contact with the face should be harmless to the skin. (b) The filter media shall be harmless to humans. (c) The material used shall have sufficient strength to withstand breakage or deformation during its normal service life. Structure: (a) should not be susceptible to structural damage and the design, composition and installation of the components should not pose any hazard to the user. (b) The headband should be designed to be adjustable and easy to wear and remove, should securely fasten the mask to the face and should be worn without visible compression or pain, and the headband design of the replaceable half mask and full mask should be replaceable. (c) Should have as small a dead space and a large field of view as possible. (d) The disposable mask shall be constructed to ensure a close fit to the face and shall be free from deformation during its service life. | Struc | 300 | | | |
| 2 | External Inspection | <u>-</u> | The surface shall be free from breakage, deformation and other obvious defects, the material and construction of the components shall be able to withstand normal conditions of use and the temperature, humidity and mechanical shocks that may be encountered, the headband shall be adjustable and the headband design of the replaceable mask shall be replaceable. Parts should not be dislodged, damaged or deformed after pre-treatment with temperature and humidity. | Meet t | Qualifie | | | |
| 3 | Bri | | CO V | | Sample No. | 31 | MV | |
| | Respiratory resistance | | | Total inspir | Unproce | 1# | 109.3 | |
| | | | Total inhalation resistance should not | atory resist ance | Pretreat | 2 [#] 5 [#] | 110.3 | |
| | | Pa | exceed 350 and total exhalation resistance should not exceed 250 for each sample. | | | 6# | 112.3 | 0 115 |
| | | 1 a | | | ment | # | Qualified | |
| | | MY | | Total exhal | Unproce | 3# | 79.6 | |
| | | W | | ation | ssed | 4# | 83.4 | |
| | ~ (00) | | (00) | resist | Pretreat | 7 [#] | 84.5 | |
| | 5710 | | | ance | ment | 8 | 87.2 |) |
| 4 | Vision | | Lower field of view ≥60° | | Qualifie | | | |

| Serial | Test Items | Unit | D | W | | Page 3 o | J1 J | Individ |
|--------|-------------------|---------|---|---|-----------------|-----------|------|---------------------|
| Jorial | Test Hellis Uni | | Requirement | Results | | | | Individu Judgmei |
| | (32) | | (6.2) | Samp | le No. | M | V | - Guagino |
| | 3 12 | | ~ 100° | | 9# | Initial | 99.0 | |
| | 2 | | 000 | | 9" | Loaded | 98.5 | |
| | | | WT. | | 10# | Initial | 98.8 | |
| | | | | | 10# | Loaded | 98.3 | |
| | | | | 1 | 11# | Initial | 99.0 | |
| | | 9 | | | 11 | Loaded | 98.4 | 0 |
| | - | UR | | Unproc | 12# | Initial | 99.1 | - AB |
| | (00) | 7 | (0) | | 12 | Loaded | 98.5 | 0) 100 |
| | 12 W | | | | 13# | Initial | 98.7 | 5 |
| - | (32) | | | | 13" | Loaded | 98.2 | |
| 64 | 3 100 | | 838 | | 14# | Initial | 99.0 | |
| 6.4 | | | 120 | | 14" | Loaded | 98.4 | Qualified |
| | Filtration | | | | 15# | Initial | 99.1 | |
| | efficiency/% | | | | | Loaded | 98.5 | |
| 5 | (NaCl particulate | | KN95: ≥95.0 | | 16 [#] | Initial | 98.9 | |
| | matter) | 5 | ~ 8 | 20 | | Loaded | 98.3 | |
| | \sim | | | | | Initial | 98.8 | |
| | ~ (00) | | (.4) | | | Loaded | 98.2 | |
| 624 | 2100 | | | | 18# | Initial | 99.0 | |
| 5 | 1100 | | 000 | | | Loaded | 98.4 | |
| CV | 30 | | 62 | | 19# | Initial | 98.6 | |
| 1 | | | 620 | | 12 | Loaded | 98.1 | |
| | | | | | 20# | Initial | 99.0 | |
| | | | | | 20 | Loaded | 98.4 | |
| | | | | Pretreat | 21# | Initial | 98.5 | |
| | | | CO | ment | 21 | Loaded | 98.0 | 25 |
| | 50 | | | 1 | 22# | Initial | 98.0 | |
| | 000 | | ~ (0.7 | | 22 | Loaded | 97.4 | |
| | (0) | | | | 23# | Initial . | 98.9 | |
| | 100 | | (400 | | | Loaded | 98.2 | |
| 6 | Headband | _ 80 | Each headband, buckle and other adjusting parts of the mask should not slip or break when subjected to a 10N tension for 10s. | Not pro Each he adjustin not slip subjecte After pr Each he adjustin not slip subjectes | Qualified | | | |
| 7 | Dead Space | | CO ₂ volume fraction should be no greater than 1%. | CO ₂ | Qualified | | | |

| Serial | Test Items | Unit | | Requirement | | | Page 4 of 5 Results | Individua |
|--------|--|-----------|---|--|---|-----------------|--|-------------|
| 8 | Breath valve cover | | The exhalation valve cover is subjected to 10N axial tension for 10s, there should be no slippage, fracture and deformation. Without this component, this item will not be evaluated | | | | Judgmen | |
| 9 | Breathing valve air tightness | 30 | the foll pumpir mL/min is less t expirate | ample shall be free from one of owing conditions.(a) When the ng flow rate has reached 500 n, the system negative pressure than 1180 Pa.(b) The time for the ory valve to return to heric pressure is less than 20s. | Without this component, this item will not be evaluated | | | To constant |
| 6 | 100 | | Filter media level | Based on Til of each action (i. e. 10 persons and 5 actions). TIL of at least 46 actions in 50 | | 26 [#] | Thora one 47 actions and | |
| 64 | 0 | | | actions | Unproc essed | 28# | Thomas 47 41 14 | |
| | | | | <11 | | 29# | 77 45 4 44 | |
| | Leakage/% (Total Inward Leakage of Disposable Mask) | | | | | 30# | There are 47actionss with a TIL less than 11. | |
| | | D. | | | Pretreat ment | 31# | There are 47 actionss with a TIL less than 11. | |
| | | | | | | 32# | There are 47 actionss with a TIL less than 11. | |
| | | | | | | 33# | There are 47actionss with a TIL less than 11. | |
| T. | | | | | | 34# | There are 47 actionss with a TIL less than 11. | |
| 10 | | | | | | 35# | There are 47 actionss with a TIL less than 11. | |
| | | | P:14 | Total Inward Leakage for at least 8 out of 10 subjects when evaluated on a total person basis | | 26# | There were 9 people with a TIL less than 8. | |
| | | 2 | Filter media | | Unproc | 27# | There were 9 people with a TIL less than 8. | |
| | | | level | | | 28# | There were 9 people with a TIL less than 8. | |
| 6 | | | KN95 | <8 | | 29# | There were 9 people with a TIL less than 8. | |
| | | | | | | 30# | There were 9 people with a TIL less than 8. | |
| | 0 | | | | | 31# | There were 9 people with a TIL less than 8. | |
| | | | | | | 32# | There were 9 people with a TIL less than 8. | |
| | | 8 | | | Pretreat | 22# | There were 9 people with | |
| | | | 1 | | ment | 2.4# | a TIL less than 8. There were 9 people with a TIL less than 8. | |
| | | | | m (0°) | | 25# | There were 9 people with a TIL less than 8. | |

STFWT20207229

| | 311 W 120201229 | A A PC | | Page 5 of 5 | | | |
|--------|-----------------|--------|---|--|------------------------|--|--|
| Serial | Test Items | Unit | Requirement | Results | Individual Judgment | | |
| 11 | Flammability | _ | The parts exposed to the flame should not burn after being removed from the flame; if they do, the renewal time should not exceed 5s. | Not processed: 1*, 2*The parts of the specimen exposed to the flame did not burn after being removed from the flame. After pretreatment: 3*, 4*The parts of the specimen exposed to the flame did not burn after being removed from the flame. | | | |

Test sample picture



-The following blank-

注 意 事 项

- 1、检验检测报告无"检验检测报告专用章"或检验检测单位公章无效。
- 2、复制检验检测报告未重新加盖"检验检测报告专用章"或检验检测单位公章无效。
- 3、检验检测报告无主检、审核、批准人签字无效。
- 4、检验检测报告涂改无效。

Points For Attention

- 1. This Report Is Invalid If Without "The Text Report Special Seal" Or The Official Seal Of The Institute.
- 2. The Reproduction Is Invalid If Without Being Confirmed By "The Text Report Special Seal" Or The Official Seal Of The Institute.
- 3. This Report Is Invalid If Without Signature Of The Major Tester And The Examiner And The Approver.
- 4. This Report Is Invalid If In Any Form By Any Means Altered.

检验检测机构地址: 江苏省泰州市高港区临港经济园临港大道166号

The Institute Add: Lingang Road 166, Lingang Economic Park, Gaogang, Taizhou. Jiangsu

检验检测机构监督电话: 0523-86989901

The Institute Complain Tel:0523-86989901

检验检测机构业务电话: 0523-86989959

The Institute Businese Tel:0523-86989959

检验检测机构传真: 0523-86989939

The Institute Fax:0523-86989939

检验检测机构邮编: 225300

The Institute Post:225300

检验检测机构网址: www.jstfzx.com

The Institute Web:www.jstfzx.com

检验检测机构邮箱: 1735889887@qq.com

The Institute E-mail:1735889887@qq.com