



中国认可
国际互认
检测
TESTING
CNAS L5772

TEST REPORT

检测报告



Commissioned by:

委 托 方

Ningbo Fanwei Arts & Crafts Co., Ltd.
宁波市泛威工艺品有限公司

Factory:

生 产 单 位

Ningbo Fanwei Arts & Crafts Co., Ltd.
宁波市泛威工艺品有限公司

Name of Sample:

样 品 名 称

KN95 protective mask (non-medical)
KN95 防护口罩（非医用）

Model:

型 号

KW901

Testing Purpose:

检 测 类 别

Commissioned
委托检测

Testing Lab:

检 测 单 位

Precise Testing & Certification (Guangdong)
Co., Ltd. (PTC)
精准通检测认证（广东）有限公司

Address:

地 址

Building 1, No. 6, Tongxin Road, Dongcheng
Street, Dongguan, Guangdong
广东省东莞市东城街道同新路6号1栋





TEST REPORT

检测报告

Report No.报告编号:PTC22032700102C-CEN01

Issue Date 报告日期: Apr.02,2022

Page 2 of 6

Name of sample 产品名称	KN95 protective mask (non-medical) KN95 防护口罩 (非医用)	Model 型号	KW901
		Specification 规格	15.5cmX11cm
Grade of sample 样品等级	/	Sample No.样品编号	/
Testing purpose 检测类别	Commission 委托检测	Batch No.生产批号	/
Quantity of sample 样品数量	/	Package 包装	Sealed 密封包装
Sample Received Date 收样日期	Mar.29,2022 2022 年 03 月 29 日	Completed Date 测试日期	Apr.02,2022 2022 年 04 月 02 日
Testing Environment 检测环境说明	As requirement of testing standard 按标准要求		
Commissioned by& Address 委托单位及地址	Ningbo Fanwei Arts & Crafts Co., Ltd. No. 63, Jiangnan East Road, Xiaogang Street, Beilun District, Ningbo City, Zhejiang Province 宁波市泛威工艺品有限公司 浙江省宁波市北仑区小港街道江南东路 63 号		
Name & Address of Factory 生产单位及地址	Ningbo Fanwei Arts & Crafts Co., Ltd. No. 63, Jiangnan East Road, Xiaogang Street, Beilun District, Ningbo City, Zhejiang Province 宁波市泛威工艺品有限公司 浙江省宁波市北仑区小港街道江南东路 63 号		
Testing Reference 检测依据	GB 2626-2019 Respiratory protection - Non-powered air-purifying particle respirator. GB 2626-2019 《呼吸防护 自吸过滤式防颗粒物呼吸器》		
Testing Item 检测项目	Filter efficiency(as received),Inhalation resistance(as received), Exhalation resistance(as received),Headband(as received) 过滤效率(未预处理)、吸气阻力(未预处理)、呼气阻力(未预处理)、 头带(未预处理)		
Remark 检测结论	Test results are attached as below 见检测结果		
Note 备注	None 无		

编制: 万芷坤

审核: 张丽红

批准: 李秋萍

万芷坤

张丽红



李秋萍

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission of the company.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)

Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 [http:// www.ptc-testing.com](http://www.ptc-testing.com)



TEST REPORT

检测报告

Report No.报告编号:PTC22032700102C-CEN01

Issue Date 报告日期: Apr.02,2022

Page 3 of 6

No. 序号	Testing Item/Unit 检测项目/单位	Testing Method 检测方法	Requirements 要求	Testing Result 检测结果		Conclusi on 结论
1	Filter efficiency 过滤效率 (%)	GB 2626-2019 Clause 6.3 GB 2626-2019 条款 6.3	KN90: $\geq 90.0\%$ KN95: $\geq 95.0\%$ KN100: $\geq 99.97\%$	As received (NaCl) 未预处理 (氯化钠)	99.26	KN95 Pass KN95 合格
					99.59	
					98.79	
					99.10	
					99.67	
					99.67	
					99.59	
					99.67	
					99.51	
					99.75	
					99.59	
					99.51	
					99.43	
2	Inhalation resistance 吸气阻力 (Pa)	GB 2626-2019 Clause 6.5 GB 2626-2019 条款 6.5	See Annex 1 见附表 1	As received 未预处理	172.15	KN95 Pass KN95 合格
					182.10	
3	Exhalation resistance 呼气阻力 (Pa)	GB 2626-2019 Clause 6.6 GB 2626-2019 条款 6.6	See Annex 1 见附表 1	As received 未预处理	164.73	KN95 Pass KN95 合格
					167.76	

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission of the company.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)

Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 [http:// www.ptc-testing.com](http://www.ptc-testing.com)



TEST REPORT

检测报告

Report No.报告编号:PTC22032700102C-CEN01

Issue Date 报告日期: Apr.02,2022

Page 4 of 6

4	Headband 头带	GB 2626-2019 Clause 6.11 GB 2626-2019 条款 6.11	Disposable facepiece: 10N continuous 10s No slippage or breakage 呼吸器的每条头带、带扣及其他 调节部件在承受如下规定的 拉力时,不应出现滑脱或断裂。 1) 随弃式面罩: 10N,持续 10s 2) 可更换式半面罩: 50N,持续 10s 3) 全面罩: 150N,持续 10s.	As received 未预处理	Comply 符合	Pass 合格
---	----------------	--	---	---------------------	--------------	------------

Annex 1: limits of Inhalation/Exhalation Resistance(Pa)

附表 1: 吸气/呼气阻力限值 (Pa)

Type of facepiece 面罩类别	KN90 and KP90	KN95 and KP95	KN100 and KP100	Exhalation resistance / Pa 呼气阻力 /Pa
Disposable facepiece, without exhalation valve 随弃式面罩, 无呼气阀	≤170	≤210	≤250	Same as inhalation resistance 与吸气阻力一样
Disposable facepiece, with exhalation valve 随弃式面罩, 有呼气阀	≤210	≤250	≤300	≤150
Replaceable half facepiece and full facepiece including filter element 包括过滤元件在内的可 更换式半面罩和全面罩	≤250	≤300	≤350	

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission of the company.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)

Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 [http:// www.ptc-testing.com](http://www.ptc-testing.com)



TEST REPORT

检测报告



Report No. 报告编号: PTC22032700102C-CEN01

Issue Date 报告日期: Apr.02,2022

Page 5 of 6

Photo(s) of sample:

样品照片:



This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission of the company.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)

Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 [http:// www.ptc-testing.com](http://www.ptc-testing.com)




TEST REPORT

检测报告

Report No. 报告编号: PTC22032700102C-CEN01

Issue Date 报告日期: Apr.02,2022

Page 6 of 6



End of Report

报告完

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission of the company.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)

Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 [http:// www.ptc-testing.com](http://www.ptc-testing.com)