

3M Valley Plant
600 East Meigs Street
PO Box 368
Valley, NE 68064-0368
402/359-2131



Uncorded Patient Plates
CERTIFICATE OF ANALYSIS

Invoice #: NA
11 Digit #/SAP# 70201187666 / #N/A
Product 9130 3M UNIV ELECTROSURGICAL PLATE,STD,UNCORDED
Batch #: 202509MF

Expiration Date: September 2025

Manufactured Date: SEPTEMBER 2022


Tests/Equipment: TS-633 / Vector Tester
TS-05-011255 / Tensile/Peel Tester

Testing Results:

TEST	SPECIFICATIONS	RESULTS
Electricals:		
Impedance	UISL= 0.7 ohms/mil	0.168 OHMS/mil AVG.
Seal Integrity:	No open seals	PASS
Visuals:		NO DEFECTS
Count/Packaging:		NO DEFECTS

All test results comply with the current 3M specifications and product is produced under GMP guidelines.

Author:

Signed: 	Title: Cert. Coordinator	Date: 5-3-2023
---	-----------------------------	-------------------