



Draeger Colombia
Carrera 11A No 9B-50 Oficinas 603-604
Bogotá, Colombia.
pedidoscolombia@draeger.com

Permiso de comercialización:
INVIMA 2019EBC-0019909

VentStar MRI (N) 300

Dräger

REF MP04802 RI 03
LOT 147814

2020-09-30 2023-09-30



LATEX



(01)04048675634904
(11)200030
(17)200800
(10)147814
(93)MP04802-03

-20°C
-4°F

60°C
140°F

5% 95%

500 hPa
7,3 psi

1200 hPa
17,4 psi

CE 0123



1
For USA: Rx only

C@60mbar<0,6ml/mbar

R@30L/min<0,2mbar

Quality by Dräger

www.draeger.com

Made in Canada

MP04802-00000002