





TEST REPORT

(Electronic version)

扫码下载报告



VERIFICATION WEBSITE: www.gttc.net.cn VERIFICATION CODE: VIHP-1430-34

ISSUE DATE: 2021-01-27

APPLICANT: GUANGZHOU HWK MEDICAL TECHNOLOGY CO., LTD

ADDRESS: FLOOR 3. BUILDING 1. NO. 15. JINSHI AVENUE. SHILING TOWN. HUADU

DISTRICT. GUANGZHOU. GUANGDONG. CHINA

INFORMATION CONFIRMED BY APPLICANT:

KN95 DISPOSABLE PROTECTIVE MASK (NON-MEDICAL)

QUANTITY: 40 PIECES MODEL: KN95-HWK-01

No: 210021052

MANUFACTURE'S NAME: GUANGZHOU HWK MEDICAL TECHNOLOGY CO., LTD

STANDARD ADOPTED:

GB 2626-2019 (Respiratory protection-Non-powered air-powered air-purifying particle respirator)

DATE RECEIVED/DATE TEST STARTED: 2021-01-18

CONCLUSION:

APPEARANCE QUALITY[2 PIECES]

M

FILTRATION EFFICIENCY TO NaCl PARTICULATE MATTER

M

NOTE: "M"-MEET THE STANDARD'S REQUIREMENT "F"-FAIL TO MEET THE STANDARD'S REQUIREMENT "---"-NO COMMENT

REMARK:

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 210021051.

ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).

COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.

THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P.R.CHINA.

APPROVED BY:

ZiShan Guo SENIOR ENGINEER

郭子山









TEST REPORT

(Electronic version)

No:210021052





总部:广州市番禺区珠江路1号 花都实验室:广州市花都区狮岭镇旗岭河滨西路1号 电话:020-61994598/61994599 电话:020-37721161/66348638







TEST REPORT

(Electronic version)

No:210021052

APPEARANCE QUALITY[2 PIECES]

M

(GB 2626-2019 6.1)

REQUIREMENT

PASS

ACCORDING TO THE CLAUSE 5.2 OF THE

PRODUCT STANDARD (GB 2626-2019)

FILTRATION EFFICIENCY TO NaC1 PARTICULATE MATTER(%)

M

(GB 2626-2019 6.3, AIR FLOW: 85L/min, AEROSOL: NaC1, AEROSOL CONCENTRATION: 15mg/m³,

TEMP: 22.4°C RH: 34.2%)

REQUIREMENT

FILTRATION EFFICIENCY:

 ≥ 95.0 (KN95)

(GB 2626-2019)

FILTRATION EFFICIENCY: UNTREATED SAMPLE

1# 99.982 99.990

2#

3# 99.878

4# 99.973

99.873 5# 99.991 6#

7# 99.765

8# 99.884

9# 99.857

10# 99.965

11# 99.837 12# 99.961

13# 99.875

14# 99.991

15# 99.952

CONDITIONING TREATED

1# 99.924

2# 99.991

99.992 3#

4# 99.936

5# 99.947



----End of Report-----